



Industrie Service

CERTIFICATE

The Testing and Calibration Station
of TÜV SÜD Industrie Service GmbH

confirms to the company

Dreyer + Timm GmbH
Spitzen 1
27389 Fintel

based on the positive test results of the tested temperature recorder,

scomsens L

according to the test-report No. **1101 TS** from the 12th of August 2013, that the before named temperature recorder fulfils the requirements of the standard EN 12830 (Status Oct. 1999).

Accuracy class: **1**
Environmental conditions **C, D**
Application range: **S,T**

The manufacturer is allowed to mark the temperature recorders with the following test-report number:

ATP-MUC 1101 TS

This certificate is valid from **12.08.2013**

This certificate is valid to: **11.08.2018**

Munich, the **12th of August 2013**

Center of Competence
Refrigeration and air-conditioning

Test engineer

Andreas Klotz



Thorsten Bennert